

# From Eye to Insight



Leica Microsystems (Schweiz) AG | Max Schmidheiny-Strasse 201 | 9435 Heerbrugg (Switzerland)

To whom it may concern,

30<sup>th</sup> June 2023


## **Antibacterial Activity Evaluation Test declaration**

Proof of antibacterial evaluation test according to JIS Z 2801:2010/A1:2012

Coating of below mentioned products pass antibacterial activity evaluation test according to JIS Z 2801:2010/A1:2012 with a value of 3.78 for the antibacterial activity (Pass criteria is >2.0) for Staphylococcus aureus (ATCC 6538P) and Escherichia coli (ATCC 8739) The test rapport confirms an effectivity of >3.78 which means >99.99% killing ratio.

Order Nr.	Product
10 450 981	Ivesta 3 (Integrated camera)
10 450 982	Ivesta 3 (C-mount)
10 450 983	Ivesta 3

The antibacterial activity evaluation test according to JIS Z 2801:2010/A1:2012 was done by TÜV Süd Singapore, test report numbers 7191259112-CHM21-01-RC\_CR1 and 7191259113-CHM21-01-RC\_CR1.

DocuSigned by:  
  
12B020BA97F540E...

Dr. Jochen Langer  
Senior Manager Global Product Compliance